



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 14, 2012 INITIALS: Dec OWNER: SNOHOMISH COUNTY  
DISTRICT: ADDRESS:  
LOCATION: FM SHOP COUNTY:  
MAKE: GEC TYPE: PORT. WHEEL SCALE ACCURACY CLASS: IIII  
MODEL: MD-400 ID. NO: 25199 CAPACITY: 20000 lbs. SECTIONS:  
MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: DIAL PLATFORM 11"X12"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	9960	-40				
CENTER	15000	14920	-80				
CENTER	20000	19920	-80				


Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration",  
Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy

Signed by: David Cromer

Date: June 19, 2012

Print name: DAVID CROMER



State of Washington

County of Thurston

Signed and sworn to (or affirmed before me on June 19, 2012

Patricia A. Geppert

(Signature of notary public)

Patricia A. Geppert

Printed Name of Notary

My appointment expires Sept. 4, 2015